



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 26.09.2023 Calibration valid: 25.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/S, GOVT.PRIMARY HEALTH CENTRE, SAMUDRAN, SALAM,
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i THERMOSCIENTIFIC	THERMO SCIENTIFIC
ii FINNPIPETTE F3	FINNE PIPETTE
iii Value	100µl-1000µl
iv Serial No	RW11613

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.13	1000.13
2	100.03	500.13	1000.04
3	100.02	500.01	1000.07
4	100.04	500.02	1000.08
5	100.00	500.01	1000.09
6	100.08	500.07	1000.02

For Heart Beat Medical Systems,

  
(Authorized signatory)

