

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	Govt. UPHC, Therkuvasal	<b>Details About Robust Analysis</b>								
<b>Lab No</b>	10819	<b>Detail About Monthly Summary</b>								
<b>Month</b>	July	<b>Detail about SDI</b>								
<b>Year</b>	2023									
<b>Constituent Group</b>	Chemistry III									
		Click on the analyte to view Graphical Data				<a href="#">All Analyser Result</a>		<a href="#">Print</a>		
		Date of Result Entered : 20/07/2023								
		Date of Report Published : 07/08/2023								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	110 mg/dL	0.11	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	7.3 mg/dL	4.2	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	145 mg/dL	6.87	0.22
		<b>SDI Range</b>		<b>Interpretation</b>						
		Within -1.00 to +1.00		Excellent.						
		Within ±1.01 to ±2.00		Good.						
		Within ±2.01 to ±2.99		Accept with caution. Warning Signal.						
		Beyond ±3.0		Unacceptable performance. Action Signal.						

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>