

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	Govt. UPHC, Therkuvasal	<b>Details About Robust Analysis</b>								
<b>Lab No</b>	10819	<b>Detail About Monthly Summary</b>								
<b>Month</b>	September	<b>Detail about SDI</b>								
<b>Year</b>	2023									
<b>Constituent Group</b>	Chemistry III									
Click on the analyte to view Graphical Data						<a href="#">All Analyser Result</a>		<a href="#">Print</a>		
Date of Result Entered : 20/09/2023										
Date of Report Published : 05/10/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	311 mg/dL	0.01	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	110 mg/dL	-0.84	0.20
<b>SDI Range</b>			<b>Interpretation</b>							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>