



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 26.09.2023 Calibration valid: 25.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/S, GOVT.PRIMARY HEALTH CENTRE, GOODAMALAI, ATTUR HUD, GENGAVALLI BLOCK.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNPIPETTE F3
iii Value	10 µl -100 µl
iv Serial No	RW07834

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.05	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.06
4	10.09	50.08	100.08
5	10.00	50.09	100.09
6	10.08	50.08	100.11

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 26.09.2023 Calibration valid: 25.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/S, GOVT.PRIMARY HEALTH CENTRE, GOODAMALAI, ATTUR HUD, GENGAVALLI BLOCK.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i	THERMOSCIENTIFIC -
ii	FINNPIPETTE F3 -
iii	Value 100µl-1000µl
iv	Serial No PHC/GM/PI/001

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.05	500.11	1000.12
2	100.06	500.12	1000.03
3	100.02	500.01	1000.07
4	100.04	500.02	1000.08
5	100.00	500.00	1000.09
6	100.08	500.07	1000.02

For Heart Beat Medical Systems,


(Authorized signatory)





CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 26.09.2023 Calibration valid: 25.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/S, GOVT.PRIMARY HEALTH CENTRE, GOODAMALAI, ATTUR HUD, GENGAVALLI BLOCK.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	MICROLUX
ii Model	-
iii Value	10 µl
iv Serial No	PHC/GM/PI/002

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)
	10µl
1	10.02
2	10.02
3	10.05
4	10.04
5	10.01
6	10.03

For Heart Beat Medical Systems.

(Authorized signatory)

