



**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24		Calibration Date : 26.09.2023
		Calibration valid: 25.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/S, GOVT.PRIMARY HEALTH CENTRE, GOODAMALAI, ATTUR HUD, GENGAVALLI BLOCK.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
<b>i</b>	Manufactured by	B.D INSTRUMENTATION
<b>ii</b>	Model	BDI-152
<b>iii</b>	Serial No	PHC/GM/CENT/001

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1200	0
3	1500	1501	1
4	2000	2001	1

Acceptable deviation is ± 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016



GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662



No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



**Testy equipment used**

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer ( non – contact ) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

**Remarks**

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :

  
