

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMASAPURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.

E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9
Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPHC, MAMASAPURAM.

Scope:

Equipment Name	: CENTRIFUGE	Asset no	: PHC/MSP-LAB/001
Make	: REMI	Calibration Done at	: customer place
Model	: C-854/4	Calibration Date	: 29/08/2023
Sl no	: ADLC-689	Next call due on	: 28/08/2024

<u>Sl. No.</u>	<u>DUC reading(RPM)</u>	<u>Standard Reading (RPM)</u>	<u>Correction (RPM)</u>
01	Position 1	2010	+0
02	Position 2	2140	+0
03	Position 3	2460	+0
04	Position 4	3450	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

1. Digital Multimeter-Mastech-MS830L
2. Digital Taco meter-DT-2234C/Max tech

Note:

1. Calibration points were selected as per customer request.
2. Each reading was taken by repeatability.
3. Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems

Authorized signatory

