



CALIBRATION CERTIFICATE

Page No:1 of 1

CERTIFICATE NO:	SBS/CL/22/11745	MECHANICAL	
Issue Date	22-10-2022		
SRF No & Date	SRF/22/00250-0019 & 20-10-2022		
Receipt Date	20-10-2022		
Calibration Date	20-10-2022		
Calibration Due	19-10-2023		

Customer Name & Address
 GOVERNMENT COMMUNITY HEALTH CENTRE,
 NATTARAMMAN KOVIL STREET,ERAL-628801.

Details of Device Under Calibration (DUC)			
Description	CENTRIFUGE	Make & Model	LABOTECH & BDI-152
Range	4000 RPM	Sr. No	2011922
Resolution	1 RPM	Identification No	
DUC Condition	Satisfactory	Location	LAB


Environmental Conditions & Standard Operating Procedure Details			
Environmental Details	Temperature: 25.3°C	Relative Humidity	54% Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details			
S.No	Description	Make/ SI No:	Certificate No
1	Digital Tachometer	LINE SEIKI / 175-0034V	TMS/21/257 02
			Validity 11-11-2022

CALIBRATION RESULTS				
S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	3500	3497	-3	4.2

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor $k=2.00$.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 M.BALAJI



Quality Manager
 (D.VETRI SELVI)

Chief Executive