



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name	Govt. PHC, Mamsapuram	Lab No	13162
Constituent Group	Chemistry III	Date of Result Entered :	20/09/2023
PT item	Lyophilized human serum based	Date of Report Published :	05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyte Volume	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robert (semi automatic Fully Automatic)	1870	310.86	4.82	14.98	319.8 mg/dL	0.60	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robert (semi automatic Fully Automatic)	1852	1.29	8.14	0.10	1.28 mg/dL	-0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robert (semi automatic Fully Automatic)	1783	113.62	3.80	4.32	105.4 mg/dL	-1.90	0.20

SDI Range	Interpretation
Within ±1.0 to +1.00	Pass
Within ±1.01 to ±2.00	Caution
Within ±2.01 to ±2.99	Action. Warning Signal.
Beyond ±3.0	Unsatisfactory performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Mamsapuram
 CLINICAL LABORATORY, SRIVILLIPUTHUR, VIRUDHACHARI DISTRICT
 Mamsapuram
 Tamil Nadu - 626125

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity of the sample is passed.
 Data integrity reports is confidential
 CMC EQAS do not contract any components
 of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name	Govt. PHC, Mamsapuram	Lab No	13162
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	184.6 mg/dL	1.00	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	118.1 mg/dL	0.93	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Mamsapuram
CLINICAL LABORATORY, SRIVILLIPUTHUR, VIRUDHINAGAR
Mamsapuram
TAMILNADU 626125

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****