

View Lab Monthly Summary

Lab Name Govt. PHC, Ulkottai
 Lab No 12007
 Month April
 Year 2023
 Constituent Group Chemistry III

[Details About Robust A](#)[Detail About Monthly S](#)[Detail about SDI](#)

Click on the analyte to view
Graphical Data

All Analyser Result

Date of Result Entered .
20/04/2023

Date of Report Published
06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	109 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	118 mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#) [Modify Methods](#) [Submit Lab Results](#) [View Lab Monthly S](#)
[Yearly Summary](#) [Change Password](#) [Home](#) [Sign-out](#)