

			<u>C</u>	ALIBRA	TION C	ER1	TIFIC/	ATE_		
CERTIFICATE NO: SBS/CL/23/13176				MEDICAL DEVICES			EVICES	Page No:1 of 1		
Issue Date				07-10-2023						
SRF No & Date					SRF/23/00691-0001 & 06-10-2023					
Receipt Date					06-10-2023					
Calibration Date					06-10-2023					
Calibration Due					05-10-2024					
Custo	omer Name	& Add	iress							
GOVE	RNMENT	PRIMA	RY HEALTH C	ENTRE,						
			2,KRISHNAGIF	RI DISTRICT.		1-0				
				Details of	Device Unde	r Calibra	tion (DUC			
Dascr	intion		SEMI AUTO A	NALYZER	Make & N	Model	:	ROBONIK & PRI	ETEST TOUCH	
Description : SEMI AUTO AN Range : MULTI			Sr. No			ATCD3650321R	BK			
Resolution : MULTI			Identification No		:	NA				
	Condition		SATISFACTO	RY	Location		:	LABORATORY		
-	OONGIGON		E	nvironmental Co	nditions & C	allbratio	Procedu	e Details		
Environmental Details   Temperature:25.6							ty 52% RH			
Calibration Procedure No SBS/CP/MD						done at ONSITE				
					ference Stan	dards De	talls	Tarken day be		
S.No Description			Make/ SI	Make/ SI No:		Certificate No		Validity		
1 Electrical Safety Analyser			Rigel Medical & 44L-1059		M-230809-16-4		10-08-202			

RESULTS							
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)				
-	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)				
1	>20MΩ	92	13.92				
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)				
4	<5000µAfor B.BF.CF	172	7.5				
-		Measured values in µA	Uncertainty in % (±)				
3	Enclosure Leakage	. 221	7.2				

**ELECTRICAL SAFETY** 

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.

<500µAfor B,BF,CF

- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY





Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

[No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennal - 600 032, Tel: 044 - 2225 2087,

[E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.