



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. Block PHC, Karipatti

Lab No

7494

Constituent Group Chemistry III

Date of Result Entered :

20/09/2023

PT item Lyophilized human serum based

Date of Report Published :

05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	323 mg/dL	0.81	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Pictate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.29 mg/dL	0.00	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	109 mg/dL	-1.07	0.20
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Karipatti

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.

Karipatti

TAMILNADU636106