



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12821

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE(KANAI BLOCK), KANAI-605301,VILLUPURAM DISTRICT.	SRF No.	SRF/23/00609-0006
	SRF Date	04-10-2023
	Date of Receipt	04-10-2023
	Date of Calibration	04-10-2023
	Due Date for Calibration	03-10-2024
	Issue Date	05-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	DRAGON LAB
Range	100-1000µl	Model	NA
Resolution	5µl	Material	PVC
Serial Number	YE195AL0408360	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.2	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.93	99.94	99.92	99.95	99.96	99.94	-0.06	0.02	0.47
		99.95	99.93	99.96	99.97	99.93				
2	500	499.92	499.95	499.96	499.94	499.96	499.96	-0.04	0.02	0.47
		499.95	499.97	499.98	499.99	499.96				
3	1000	999.95	999.96	999.97	999.98	999.96	999.92	-0.08	0.09	0.47
		999.95	999.98	999.96	999.74	999.75				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,


 (Calibration Engineer)
 M.ADHIBAN



Authorised by:


 (Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.