

MAR-2023



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. Block PHC, Sathiamangalam Lab No 7275
 Constituent Group Chemistry III Date of Result Entered : 18/03/2023
 PT item Lyophilized human serum based Date of Report Published : 04/04/2023

| Sl. No. | Test Name | Method | Instrument | Result | Target | Delta | Delta 2 | Delta 3 | SDI | U |
|---------|----------------|---------------------------------------|--|--------|--------|-------|---------|-----------|-------|------|
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1878 | 91.48 | 7.46 | 6.82 | 83 mg/dL | -1.24 | 0.31 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1851 | 1.33 | 9.08 | 0.12 | 1.4 mg/dL | 0.58 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1761 | 110.93 | 5.23 | 5.80 | 116 mg/dL | 0.87 | 0.28 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. Block PHC, Sathiamangalam

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.

Sathiamangalam

TAMILNADU604153

Pamela Christudoss

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

P. S. Srinivasan