



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name	Govt. PHC, Alampalayam	Lab No	12181
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT Item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl No	Analyte	Method / Principle / Name	Analyzer Name	No of Participants	D _r	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	179 mg/dL	0.38	0.42
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	15	1.72	6.81	0.12	1.75 mg/dL	0.26	0.06
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	111 mg/dL	-0.60	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


LAB ADDRESS :

Govt. PHC, Alampalayam
THE MEDICAL OFFICER, AMMAPET
Alampalayam
TAMILNADU638501

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****


MEDICAL OFFICER,
Govt. Primary Health Centre
Alampalayam,
do (DU) - 638 502.