



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name	Govt. PHC, Alampalayam	Lab No	12181
Constituent Group	Chemistry III	Date of Result Entered :	19/04/2023
PT item	Lyophilized human serum based	Date of Report Published :	06/05/2023

Sl No	Analyte	Method - Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	114 mg/dL	-0.76	0.34
2	CREATININE III	OTHERS ( Any Other Principles / Methods )	Any Analyser (Automation / Semi Automation )	13	2.20	10.23	0.22	2.19 mg/dL	-0.04	0.12
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	111 mg/dL	-0.11	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. PHC, Alampalayam  
 THE MEDICAL OFFICER, AMMAPET  
 Alampalayam  
 TAMILNADU638501

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

**MEDICAL OFFICER.**  
 Govt. Primary Health Centr  
 Alampalayam,  
 de (Dt) - 638 502