



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



12181

20/03/2023

04/04/2023

Lab Name Govt. PHC, Alampalayam
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No
Date of Result Entered :
Date of Report Published :

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DY	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	84 mg/dL	-1.10	0.31
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	14	1.32	8.94	0.12	1.25 mg/dL	-0.59	0.06
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	105 mg/dL	-1.02	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


LAB ADDRESS :
Govt. PHC, Alampalayam
THE MEDICAL OFFICER, AMMAPET
Alampalayam
TAMILNADU638501

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****


MEDICAL OFFICER,
Govt. Primary Health Centre
Alampalayam,
Tamil Nadu - 638 502.