

CALIBRATION CERTIFICATE

Certificate Number :	MED/R/CAL-69	Date of issue:	29-Sep-23
ULR No.		Job Order/ CRF No. :	
Calibration Date	28-Sep-23	Calibration Due	27-Sep-24

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
Name	Govt Primary Health centre	Description	SEMIAUTO ANALYZER
Address	Alampalayam Erode	Range / Resolution	
		Make	Genuine Biosystem
		Model	GB235
		Serial No.	BACL1222
		Identification No	AGPHCLAB'SAA01
		Location	LAB


ENVIRONMENTAL CONDITION		MASTER EQUIPMENT DETAILS	
Temperature	23 ± 2°C	Description	MULTIMETER
Humidity	50 ± 10 % Rh	Certificate No	FCL/23/10-01
Air Pressure	1010 mbar	Identification No	MEDEX-11
		Calibration Due	7-May-24
		Traceability	National Standards

SL NO	PARTS	CLEANING	RESULT
1	Main voltage: Live to Earth	226.4 v	PASS
2	Main voltage: Neutral to Earth	1 v	PASS
3	Main voltage: Live to Earth	228.6 v	PASS

Note:

- The UUC is calibrated at Lab.
- This calibration certificate relates only to the UUC supplied.
- Calibration certificate shall not be reproduced, except in full, without written approval of medex labtech solutions.
- Results reported are valid at the time and under stated environmental conditions of calibration.
- The uncertainty of measurement is expressed at 95.45% confidence level with coverage factor k=2
- This calibration certificate is issued for scientific or industrial purpose only.
- Unless Otherwise Specified the Measurement Data Reported is "As Found" -without any adjustment.
- Next calibration due as per customer's plan and requirement.
- Decision Rule Required : No

Calibrated by


Vipin PR
Calibration Engineer

Authorised By


Jude Joseph Raj
Quality Manager



X/657/28, 2nd Floor, RP Tower, Opp. Chaldean Centre,
Pallikulam Road, Thrikkur, Kerala, India - 680001

Email sales@medexlabtech.com, service@medexlabtech.com
Mob: +91 9637588562, 8113816000