

**CALIBRATION CERTIFICATE**

<b>Certificate Number :</b>	MED/R/182	<b>Date of Issue:</b>	29-Sep-23
<b>ULR No.</b>	CC353523000000182F	<b>Job Order/ CRF No. :</b>	
<b>Calibration Date</b>	28-Sep-23	<b>Calibration Due</b>	27-Sep-24
<b>CUSTOMER INFORMATION</b>		<b>UNIT UNDER CALIBRATION</b>	
<b>Name</b>	Govt Primary Health centre	<b>Description</b>	CENTRIFUGE
<b>Address</b>	Alampalayam Erode	<b>Range / Resolution</b>	1000-4000 RPM
		<b>Make</b>	REMI
		<b>Model</b>	C852
		<b>Serial No.</b>	2BKN32067
		<b>Identification No</b>	AGPHC/LAB/CENT/01
		<b>Location</b>	LAB
<b>ENVIRONMENTAL CONDITION</b>		<b>MASTER EQUIPMENT DETAILS</b>	
<b>Temperature</b>	27 ± 2°C	<b>Description</b>	DIGITAL TACHOMETER
<b>Humidity</b>	50 ± 10 % Rh	<b>Certificate No</b>	CC338823000000124F
<b>Air Pressure</b>	991 hpa	<b>Identification No</b>	S320211
		<b>Calibration Due</b>	18-Aug-24
		<b>Traceability</b>	National Standards

Sl.No	UUC READING (RPM)	STANDARD READING (RPM)	Error (RPM)	Expanded Uncertainty ± RPM
1	1000	1008	8	3.00
2	2000	2002	2	3.00
3	3000	3008	8	3.00
4	3500	3506	6	3.00
5	4000	4008	8	3.00

**Note:**

1. The UUC is calibrated at Lab.
2. This calibration certificate relates only to the UUC supplied.
3. Calibration certificate shall not be reproduced, except in full, without written approval of medex labtech solutions.
4. Results reported are valid at the time and under stated environmental conditions of calibration.
5. The uncertainty of measurement is expressed at 95.45% confidence level with coverage factor  $k=2$
6. This calibration certificate is issued for scientific or industrial purpose only.
7. Unless Otherwise Specified the Measurement Data Reported is "As Found" -without any adjustment.
8. Next calibration due as per customer's plan and requirement.
9. Decision Rule Required : No

**Calibrated by**



Vipin PR  
Calibration Engineer



**Authorised By**



Jude Joseph Raj  
Quality Manager

X/657/28, 2nd Floor, RP Tower, Opp. Chaldean Centre,  
Pallikulam Road, Thrissur, Kerala, India - 680001

Email [sales@medexlabtech.com](mailto:sales@medexlabtech.com), [service@medexlabtech.com](mailto:service@medexlabtech.com)  
Mob: +91 9037588562, 8113816000