

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :		CERTIFICATE NO	TICS/2023/MM/SG/KAM/06		
GOVT.PRIMARY HEALTH CENTRE		EQUIPMENT NAME	MICROPIPETTE		
KALAIYARMANGALAM,		MANUFACTURE	MICROPET		
SIVAGANGAL.		MODEL	20 µL		
CALIBRATION DATE	03/10/2023	ASSET NO	KAM PH 06		
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB		
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	20	20.08	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE Calibrated By: <i>Roz</i> Authorized Sign With Seal					

