

## View Lab Monthly Summary

Lab Name Govt. PHC, Ulkottai  
 Lab No 12007  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	188 mg/dL	0.76	0.44
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	118 mg/dL	0.58	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)

