



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

| | | | |
|-------------------|-------------------------------|----------------------------|------------|
| Lab Name | Govt. PHC, Malaiyur | Lab No | 12962 |
| Constituent Group | Chemistry III | Date of Result Entered : | 23/05/2023 |
| PT item | Lyophilized human serum based | Date of Report Published : | 05/06/2023 |

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 181 mg/dL | 0.01 | 0.44 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1854 | 1.76 | 7.48 | 0.13 | 1.5 mg/dL | -1.97 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Any Analyser (Automation / Semi Automation) | 154 | 114.84 | 5.08 | 5.83 | 136 mg/dL | 3.63 | 0.94 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. PHC, Malaiyur
Mazhaiyur main road
Malaiyur
TAMILNADU604501

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****