

## **CHRISTIAN MEDICAL COLLEGE**

## **DEPARTMENT OF CLINICAL BIOCHEMISTRY**

## CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Malaiyur Lab No 12962

Constituent Group Chemistry III Date of Result Entered : 21/08/2023

PT item Lyophilized human serum based Date of Report Published : 01/09/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11
SI.NO						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	168 mg/dL	-0.84	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	2 mg/dL	2.02	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	149	113.96	4.36	4.97	137 mg/dL	4.64	0.81

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Malaiyur Mazhaiyur main road Malaiyur TAMILNADU604501

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\*\*\*\*\* End of Report \*\*\*\*\*\*\*