



LAB MONTHLY SUMMARY

Name Govt. PHC, Othakkadal
No 7096
Month September
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/09/2023

Date of Report Published: 05/10/2023

Analyte	Method/ Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	210.86	4.82	14.98	334 mg/dL	1.55	0.60
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.23 mg/dL	-0.57	0.00
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	116 mg/dL	0.55	0.20

SDI Range	Interpretation
in -1.00 to +1.00	Excellent.
in ±1.01 to ±2.00	Good.
in ±2.01 to ±2.99	Accept with caution. Warning Signal.
and ±3.0	Unacceptable performance. Action Signal.

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out