

OW 10 2019/05



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



Name	Govt. PHC, Prathabaramapuram	Lab No	12387
Constituent	Chemistry III	Date of Result Entered :	19/04/2023
Matrix	Lyophilized human serum based	Date of Report Published :	06/05/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	118 mg/dL	-0.20	0.34
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.4 mg/dL	1.37	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	107 mg/dL	-0.80	0.28

SDI Range	Interpretation
-1.00 to +1.00	Excellent.
±1.01 to ±2.00	Good.
±2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :
 Govt. PHC, Prathabaramapuram
 CLINICAL LABORATORY
 Prathabaramapuram
 VADU611111

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not subcontract any components
 ***** End of Report *****

Rehman
 MEDICAL OFFICER
 Govt. Primary Health Centre
 Prathabaramapuram