

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name Govt. PHC, Vilunthamavadi
 Lab No 12388
 Month July
 Year 2023
 Constituent Group Chemistry III

[Back to Search Page](#)

[Print](#)

Date of Result Entered : 25/07/2023

Date of Report Published : 07/08/2023

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants			Top Value	SDI	SI
						Cv	SD	SI			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10			1.76	
			Transasia / Erba	24	109.83	6.83	7.50			3.06	
			Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.67	111 mg/dL	0.27	0.28	
			Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95			1.37	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	6.34			0.12	
			Transasia / Erba	18	5.79	3.09	6.17			0.68	
			Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	6.35	5.2 mg/dL	1.06	0.02	
			Any Analyser (Automation / Semi Automation)	66	5.89	4.60	6.27			0.67	
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.64			1.39	
			Transasia / Erba	19	102.04	4.38	4.93			2.28	
			Robonik (Semi automatic / Fully Automatic)	1776	111.11	4.12	4.90			0.52	
			Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.76	113 mg/dL	0.45	0.80	

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewAllResult_new.asp?resDetId=267&noSamples=1&month=7&year=2023&groupId=94&yearGroup=100

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance, Action Signal.

Links

Monthly Address	Monthly Address	Submit Lab	View Lab Summary
Yearly Summary	Change Password	Home	

(Handwritten Signature)
 மருத்துவ அலுவலர்
 அரசு ஆரோக்கிய அமைச்சு
 விழுந்தமங்கலம்
 தஞ்சாவூர்.

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewAllResult_new.asp?resDetId=267&noSamples=1&month=7&year=2023&groupId=94&yearGroup=100