

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Vilunthamavadi
 Lab No 12388
 Month April
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	118 mg/dl	-0.29	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	6.18	2.4 mg/dl	1.37	0.91
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79	193 mg/dl	-1.72	0.86

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

- [Modify Address](#)
- [Modify Methods](#)
- [Submit Lab Results](#)
- [View Lab Monthly Summary](#)

https://home.cmcvellore.ac.in/clinql/ViewMonthlySummary/ViewResult2_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110

- [Yearly Summary](#)
- [Change Password](#)
- [Home](#)
- [Sign-out](#)

(Handwritten Signature)
 மருத்துவ பிழுவலா
 அரசு சட்டம்துறை நகராட்சி
 விழுந்தளவடி
 நகராட்சிமன்றம்.

https://home.cmcvellore.ac.in/clinql/ViewMonthlySummary/ViewResult2_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110