VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Vilunthamavadi
Lab No	12388
Month	May
Year	2023
Constituent Group	Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result | Print

Date of Result Entered : 19/06/2023

Date of Report Published : 06/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	178 mg/dL	-0.31	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.78	7.48	0.13	1.85 mg/dL	-0.83	0.91
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	118 mg/dL	0.54	0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary

Yearly Summary | Change Password | Home | Sign-out

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=285&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110 22

Scanned with OKEN Scanner