



VIEW LAB MONTHLY SUMMARY

Jan-23

Lab Name: Govt. Block PHC, Katchanakudi

Lab No: 7094

Month: January

Year: 2023

Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Details About Monthly Summary](#)

[Details about SDI](#)

Date of Result Entered : 19/01/2023

Date of Report Published : 09/02/2023

[Click on the analyte to view Graphical Data](#)

ANALYTE	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD	SE			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	0.38	7.05	122.8	0.38	0.35	
2	CREATININE III	Jaffes Kinetic - Alkaline Phosphate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	2.2 mg/dL	-0.05	0.01	
3	CHOLESTROL III	CHOD-PAP III	Arg Analyser (Automation / Semi Automation)	149	111.59	0.06	8.67	100.2	-2.01	0.98	
Interpretation											
SDI Range				Excellent.							
Within -1.00 to -1.00				Good.							
Within +1.01 to +2.00				Accept with caution. Warning Signal							
Within +2.01 to +2.99				Unacceptable performance. Action Signal.							
Beyond +3.0											

Handwritten notes in Telugu and English, including a signature and date: 09/02/2023.