

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13075

Customer Name & Address:

GOVERNMENT PRIMARY HEALTH CENTRE(THELLAR BLOCK),
MAZHAIYUR-604502,VANDAVASI TALUK,TIRUVANNAMALAI DISTRICT.

SRF No.	SRF/23/00665-0002
SRF Date	05-10-2023
Date of Receipt	05-10-2023
Date of Calibration	05-10-2023
Due Date for Calibration	04-10-2024
Issue Date	06-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW10092	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		9.91	9.92	9.93	9.92	9.91				
1	10.0	9.91	9.92	9.93	9.92	9.91	9.92	-0.08	0.01	0.47
		9.92	9.90	9.93	9.94	9.95				
2	50.0	49.95	49.96	49.97	49.98	49.99	49.95	-0.05	0.03	0.47
		49.90	49.91	49.92	49.92	49.96				
3	100.0	99.96	99.97	99.98	99.97	99.96	99.97	-0.03	0.01	0.47
		99.96	99.97	99.98	99.99	99.96				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.