



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name      Govt. UPHC, Dharapadavedu      Lab No      9392  
 Constituent      Chemistry III      Date of Result Entered :      19/08/2023  
 Group  
 PT item      Lyophilized human serum based      Date of Report Published :      01/09/2023

Sl. No.	Parameter	Method	Manufacturer	Lot No.	Concentration	CV (%)	Delta	Delta 2	Delta 3	Delta 4	Delta 5	Delta 6	Delta 7	Delta 8	Delta 9	Delta 10
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	167.6	-0.89	0.42						
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.74	-0.17	0.01						
3	CHOLESTEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112	-0.39	0.22						

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ± 2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond : ± 3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. UPHC, Dharapadavedu, Balaji Nagar, 1st street  
 BALAJI NAGAR, 1ST STREET, DHARAPADAVEDU  
 Balaji Nagar, 1st street, Dharapadavedu  
 TAMILNADU 632007

*Pamela Christudoss*

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 contactNumber: 0416-2283102

Dr. Pamela Christudoss  
 CMCEQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not subcontract any components  
 \*\*\*\*\*End of Report\*\*\*\*\*

DR. PAMELA CHRISTUDOSS  
 CMCEQAS COORDINATOR  
 CHRISTIAN MEDICAL COLLEGE, VELLORE  
 CONTACT NO: 0416-2283102