



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. UPHC, Dharapadavedu Lab No: 9302
 Constituent Group: Chemistry III Date of Result Entered: 19/07/2023
 PT Item: Lyophilized human serum based Date of Report Published: 07/08/2023

Sl. No.	Parameter	Method	Reagent	Result	Reference Range	CV (%)	Delta	Delta 2	Delta 3	
1	GLUCOSE III	GOD/POD III	Rebonick (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	110	0.11	0.23
2	CREATININE III	GOD/POD III	Rebonick (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.5mg	-0.94	0.02
3	CHOLESTEROL III	GOD/POD III	Rebonick (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Dharapadavedu, Balaji Nagar, 1st street
 BALAJI NAGAR, 1ST STREET, DHARAPADAVIDU
 Balaji Nagar, 1st street, Dharapadavedu
 TAMILNADU 632007

Pamela Christudoss

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMCEQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed
 Data in CMC EQAS reports is confidential
 CMCEQAS does not sub contract any components
 *****End of Report*****

MEDICAL OFFICER
 (Lab) Govt. UPHC, Dharapadavedu
 Balaji Nagar, 1st street, Dharapadavedu
 Tamil Nadu, 632007