



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUN 2023



Lab Name Govt. UPHC, DHARAPADAVIDU Lab No 9302  
 Constituent Group Chemistry III Date of Result Entered : 19/06/2023  
 PT item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl. No.	Parameter	Method	Instrument	Lot No.	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	315 mg/dl	1.91	0.69
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.61 mg/dl	0.19	0.01
3	CHOLETEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	114 mg/dl	0.05	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ± 2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond :± 3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. UPHC, Dharapadavedu, Balaji Nagar, 1st street  
 BALAJI NAGAR, 1ST STREET, DHARAPADAVIDU  
 Balaji Nagar, 1 st street , Dharapadavedu  
 TAMILNADU 632007

*Pamela Christudoss*

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMCEQAS Coordinator  
 Christian Medical College, Vellore

**CLINICAL OFFICER**  
 Quality Health Centre  
 Govt. UPHC, Karpadi - 632 607

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\*End of Report\*\*\*\*\*