



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PG-1024

Lab Name Govt. UPHC, Dharapadavedu Lab No 9302  
 Constituent Group Chemistry III Date of Result Entered : 19/05/2023  
 PT Item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl. No.	Analyte	Method	Analyzer Name	No. of Participants	DV	Participant		CV	SI	Y	Z
						CV	SI				
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	194 mg/dl	1.40	0.44	
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dl	-0.45	0.01	
3	CHOLESTEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	125 mg/dl	1.78	0.27	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### LAB ADDRESS :

Govt. UPHC, Dharapadavedu, Balaji Nagar, 1st street  
 BALAJI NAGAR, 1ST STREET, DHARAPADAVIDU  
 Balaji Nagar, 1 st street , Dharapadavedu  
 TAMILNADU 632007

*Pamela Christudoss*

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMCEQAS Coordinator  
 Christian Medical College, Vellore

*C. K. S. Babu*  
 MEDICAL OFFICER  
 Govt. Primary Health Centre  
 Dharapadavedu, Kalpasi - 637

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\*End of Report\*\*\*\*\*