



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC 1024

Lab Name: Govt. UPHC, DHARAPADAVIDU Lab No: 9302
 Constituent Group: Chemistry III Date of Result Entered: 08/04/2023
 PT item: Lyophilized human serum based Date of Report Published: 06/05/2023

Sl. No.	Test Name	Method	Analyzer Name	No. of Participants	CV	CV ₁	CV ₂	Target	SDI	CV ₃
1	GLUCOSE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	125 mg/dl	0.77	0.34
2	CREATININE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.31 mg/dl	0.86	0.01
3	CHOLESTEROL III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	101.3 mg/dl	-1.78	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Dharapadavedu, Balaji Nagar, 1st street
 BALAJI NAGAR, 1ST STREET, DHARAPADAVIDU
 Balaji Nagar, 1st street, Dharapadavedu
 TAMILNADU 632007

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 contactNumber: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMCEQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not subcontract any components
 *****End of Report*****

C. Sathish
 Director
 Quality Health Centre
 Thangamburam, Bangalore - 560017