

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Govt. PHC, Thanakkankulam

Lab No

12380

Constituent Group

Chemistry III

Date of Result Entered:

20/09/2023

PT item

Lyophilized human serum based

Date of Report Published :

05/10/2023

SLNo	Analyte Meth	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			No.	
		to the second of the second of				CV	SD	Your Value	e SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	320 mg/dL	0.61	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.24 mg/dL	-0.48	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	111 mg/dL	-0.61	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Thanakkankulam

CLINICAL LABORATORY

Thanakkankulam TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudos

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name

Govt PHC, Thanakkankulam

Lab No

12380

Constituent Group

Chemistry III

Date of Result Entered:

19/08/2023

PT ittem

Lyophilized human serum based

Date of Report Published:

01/09/2023

CI N	lo Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	, n
SLNo						CV	SD	Your Value	SUI	u
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	172 mg/dL	-0.40	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.84 mg/dL	0.67	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	114 mg/dL	0.04	0.22

SDI Range	Interpretation
Within -1.90 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

Govt. PHC, Thanakkankulam

CLINICAL LABORATORY

Thanakkankulam

TAMILNADU

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

https://www.cmoveliore.ac.in/clinoc/ViewMonthit/SummanutorintResult_new.acn/2mcPattd=200#.

Panela Christudos

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name

Govt. PHC, Thanakkankulam

Lab No

12380

Constituent Group

Chemistry III

Date of Result Entered:

19/07/2023

PT item

Lyophilized human serum based

Date of Report Published:

07/08/2023

		Method /		No of	07	Participants		Your		U
SI.No	Analyte	Principle Name	Analyzer Name	Participants	DV	CV	SD	Value	301	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	107.4 mg/dL	-0.32	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	119 mg/dL	1.29	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:
Govt. PHC, Thanakkankulam
CLINICAL LABORATORY
Thanakkankulam
TAMILNADU

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report ********

Primary Horbis Control Sincipal addition

1/1

anaa 8 arountd≡9



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name

Govt. PHC, Thanakkankulam

Lab No

12380

Constituent Group Chemistry III

Date of Result Entered :

19/06/2023

PT item

Lyophilized human serum based

Date of Report Published :

04/07/2023

	Analyte	Method /		Analyzer Name No of Participants DV	DV	Participants		Your	SDI	U
SI_No	Alleryte	Principle Name	Analyzer Name		CV	SD	Value	SUI	"	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	297 mg/dL	0.70	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	128 mg/dL	2.81	0.24

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS :

Govt. PHC, Thanakkankulam CLINICAL LABORATORY Thanakkankulam TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.ln Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Panela Christudos

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******

MEDICAL OFFICER
Primary Health Centre
Thanakkankulam
Madurai



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



Lab Name

Govt. PHC, Thanakkankulam

Lab No

12380

Constituent Group Chemistry III

Date of Result Entered :

19/05/2023

PT item

Lyophilized human serum based

Date of Report Published:

05/06/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your		U
SI.NO						CV	SD	Value	SUI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	151 mg/dL	-3.18	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.6 mg/dL	-1.21	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	121 mg/dL	1.10	0.27

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS :

Govt. PHC, Thanakkankulam CLINICAL LABORATORY Thanakkankulam TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudos

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report ********

MEDICAL OFFICER
Primary Health Centro
Thanakkankulam
Madural



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name

Govt. PHC, Thanakkankulam

Lab No

12380

Constituent Group

Chemistry III

Date of Result Entered:

19/04/2023

PT Item

Lyophilized human serum based

Date of Report Published:

06/05/2023

		Method /		No of	DW	Partic	ipants	Your	SDI	IJ	
SI No	Analyte	Principle Name	Analyzer Name	Participants	DV	CA	so	Value			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	122 mg/dL	0.35	0.34	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.04 mg/dL	-0.69	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	104 mg/dL	-1.31	0.28	

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance, Action Signal,					

LAB ADDRESS : Govt. PHC, Thanakkankulam **CLINICAL LABORATORY** Thanakkankulam **TAMILNADU**

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudos

Dr. Pamela Christudoss **CMC EQAS Coordinator** Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report ******

> MEDICAL OFFICER Primary Health Centre Thanakkankulam Madurai