

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE NERKUPPAL, SIVAGANGAI.		CERTIFICATE NO	TICS/2023/MM/SG/NK/02
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	LABOTECH
		MODEL	BDI 152
CALIBRATION DATE	03/10/2023	ASSET NO	NKPH-002
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 2	2145	± 5	PASS
		Speed 4	2798		

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE

Calibrated By :



Authorized Sign With Seal






CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE NERKUPPAL, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/MM/SG/NK/03
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE	THERMO SCIENTIFIC
		MODEL	10-100µL FINNPIPETTE F3
CALIBRATION DATE	03/10/2023	SERIAL NO	PW01433
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	9.82	± 1	PASS
		100	999.91	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Authorized Sign With Seal

Calibrated By :

[Signature]



[Signature]



CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS :		CERTIFICATE NO	TICS/2023/MM/SG/NK/04
GOVT.PRIMARY HEALTH CENTRE		EQUIPMENT NAME	MICROPIPETTE
NERKUPPAL,		MANUFACTURE	THERMO SCIENTIFIC
SIVAGANGAI.		MODEL	100-1000µL FINNPIPETTE F3
CALIBRATION DATE	03/10/2023	SERIAL NO	RW02704
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No


TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024

TEST REPORT					Temperature 25°C	Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS	
VOLUME	µL	100	100.12	± 1	PASS	
		1000	999.39	± 1	PASS	

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE
 Calibrated By:  Authorized Sign With Seal 



CALIBRATION / TESTING REPORT				
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE NERKUPPAL, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/MM/SG/NK/001	
		EQUIPMENT NAME	SEMI AUTO ANALYZER	
		MANUFACTURE	ROBNIK	
		MODEL	PRIETEST TOUCH	
CALIBRATION DATE	03/10/2023	SERIAL NO	ATCDO480321RBK	
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
PARAMETER	UNITS	VALUE	LIMITS	STATUS
Main Voltage L1-L2	V	230.5	240±10	PASS
Main Voltage L1-G(PE)	V	229.1	240±10	PASS
Main Voltage L2-G(PE)	V	0.8	<5	PASS
Load Current	A	0.404	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.19	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA, per IEC 60601-2-4				
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By: 		Authorized Sign With Seal 