

CA	LIBRATIO	ON CERT	IFICAT	E	
CERTIFICATE NO: SBS/CL/23/13676		MEDICAL DEVICES		CES P	Page No:1 of 1
Issue Date		07-10-2023		020	age No. 1 of 1
SRF No & Date		SRF/23/00794-0004 & 07-10-2023			
Receipt Date		07-10-2023			
Calibration Date		07-10-2023			
Calibration Due		06-10-2024			
Customer Name & Address					
GOVERNMENT PRIMARY HEALTH CEN	TRE,				
SENGURICHI-606115,KALLAKURICHI DI	STRICT.				
	Details of Devi	ce Under Calibrat	ion (DUC)		
Description ELECTRICAL SAI	FETY		, ,		
(MICROSCOPE)		Make & Model : LABOMED & NA			
Resolution MULTI		Sr. No : 13066075			
DUC Condition SATISFACTORY		Identification No : NA			
STATE TO TOTAL	L.	Location : LABORATORY  tions & Calibration Procedure Details			
nvironmental Details Temperatu	re:25.6°C				
Calibration Procedure No SBS/CP/MD/29		Relative Humidity Calibration done at		52% RH	
jeseret mi				ONSITE	
No Description	Make/ SI No:	ce Standards Deta			
Electrical Safety Analyser Rigel Medical			Certificate No Validity		Validity
		M-230809-16		-4	10-08-202
	ELEC	TRICAL SAFETY			
ESULTS					
S.no SPECIFICATION		MEASURED VALUES		EXPANDED UNCERTAINTY (±)	
1 Inculation Dealette		-			
>20MΩ		easured values in MΩ 88		Uncertainty in % (±)	
Earth Leakage M		leasured values in µA		13.92	
<5000µAfor B,BF,CF		169		Uncertainty in % (±)	

REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.

Enclosure Leakage

<500µAfor B,BF,CF

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00. 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Measured values in µA

241

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY



Technical Manager C.SHANMUGARAJ

Chief Executive

Uncertainty in % (±)

**Authorised Signatory** 

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

மருத்துவ அலுவரை, ந்சு ஆரம்ப சுகாதார நிலையம செங்குறிச்சி - 606 115, கள்ளக்குறிச்சி மாவட்டம்