



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13673		MEDICAL DEVICES		Page No:1 of 1
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00794-0001 & 07-10-2023			
Receipt Date	07-10-2023			
Calibration Date	07-10-2023			
Calibration Due	06-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, SENGURICHI-606115, KALLAKURICHI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH	
Range	MULTI	Sr. No	AT2590413RBK	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature: 25.6 °C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-06-2024


ELECTRICAL SAFETY

S.No	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 88	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B, BF, CF	Measured values in µA 175	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500µA for B, BF, CF	Measured values in µA 235	Uncertainty in % (±) 7.2

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By


(Calibration Engineer)
K.SATHYAMOORTHY

Authorised Signatory


Technical Manager
C.SHANMUGARAJ

Chief Executive



SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.


மருத்துவ அலுவலர்,
ரகு ஆரமய சுகாதார நிலையம்,
செங்குறிச்சி - 606 115,
கள்ளக்குறிச்சி மாவட்டம்