

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Ennamangalam
 Lab No: 12179
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Year Value	SDI	V
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	285 mg/dL	-0.11	0.69
2	CREATININE III	Jaffes Kinetik - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.3 mg/dL	-1.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	112 mg/dL	-0.56	0.76

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Interpret with caution. Warming Signal.