

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Ennamangalam  
 Lab No 12179  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)


[Print](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	190 mg/dL	0.97	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	154	114.84	5.08	5.83	125 mg/dL	1.74	0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

  
**MEDICAL OFFICER**  
 Primary Health Centre,  
 Ennamangalam - 625 007

Links

[https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResul2\\_new.aspx?resDatId=265&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110](https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResul2_new.aspx?resDatId=265&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110)