



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. UPHC, Anna Hospital

Lab No: 10840

Constituent Group: Chemistry III

Date of Result Entered: 19/08/2023

PT Item: Lyophilized human serum based

Date of Report Published: 01/09/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	176 mg/dL	0.05	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
Govt. UPHC, Anna Hospital  
THE MEDICAL OFFICER, ANNA HOSPITAL, ANMAPET,  
Anna Hospital  
TAMILNADU636003

**Coordinator Contact Details:**  
Email: cllngc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Dr. Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*