



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 03.10.2023 Calibration valid : 02.04.2024
<b>I.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , ANNA HOSPITAL, SALEM CORPORATION, SALEM.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	P FACT
ii Model	-
iii Value	100µl-1000µl
iv Serial No	238680

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.01	500.05	1000.04
2	100.01	500.01	1000.04
3	100.04	500.05	1000.07
4	100.06	500.02	1000.08
5	100.02	500.05	1000.12
6	100.05	500.07	1000.04

For Heart Beat Medical Systems,

*R. R. R.*  
(Authorized Signatory)



MD Reg No: TN/CBE/MD42/00016  
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
88833 15662

No.42/1, Navaindia Road, Peelamedu,  
CBE - 04, hbmedicalsystem@gmail.com