

			CA	LIBRA	TIO	N CERT	IFIC.	<u>ATE</u>		
CERT	CERTIFICATE NO: SBS/CL/23/13868			<u> </u>	MEDICAL DEVICES			Page No:1 of		
Issue Date					15-09-2023					
SRF No & Date				SRF/23/00829-0002 & 14-09-2023						
Receipt Date				14-09-2023						
Calibr	ation Date	!			14-09	14-09-2023				
Calibr	ation Due	-			13-09	13-09-2024				
	mer Name									
GOVE	RNMENT	PRIMA	RY HEALTH CENT	ΓRE,						
KATT	UKOTTAI-	636453,	SALEM DISTRICT	.						
				Details of	Device	Under Calibrati	ion (DUC	5)		
Description : SEMI AUTO ANALYZER		YZER	Make & Model :		ROBONIK & PRIETEST TOUCH					
Range :		MULTI		Sr. No	Sr. No :		AT1400211RBK			
Resolution : MULT		MULTI		Identi	ification No	:	NA			
DUC C	DUC Condition : SATISFACTORY			Location :		LABORATORY				
			Envi	ronmental Co	nditions	& Calibration	Procedu	re Details		
Environmental Details Temperature:25.6 ° C				re:25.6 ° C	Relative Humidity		52% RH			
Calibration Procedure No SBS/CP/MD/20			D/20	Calibration done at		ONSITE				
				Ref	erence :	Standards Deta	ils			
S.No	Description Make/ SI N		No: Certifica		Certificat	e No	Validity			
1	Electrical Safety Analyser Rigel Medic		al & 44L-1059 M-230		1-230809-16-4		10-08-202			

ELECTRICAL SAFETY

RESULTS


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)	
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)	
	>20MΩ	99	13.92	
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)	
	<5000µAfor B,BF,CF	166	7.6	
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)	
	<500µAfor B,BF,CF	226	7.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

 Calibrated By

Calibration Engineer)
P.PRASANNA





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS