



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13880 Page No : 1 of 1

Customer Name & Address		SRF No.	
GOVERNMENT PRIMARY HEALTH CENTRE, MECHERI-636453,SALEM DISTRICT.		SRF/23/00831-0003	
		SRF Date	
		03-10-2023	
		Date of Receipt	
		03-10-2023	
		Date of Calibration	
03-10-2023			
Due Date for Calibration		02-10-2024	
Issue Date		03-10-2023	

Details of Unit Under Calibration			
Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW09131	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.2	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08


Standard Used						
SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319											
Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.87	9.86	9.88	9.88	9.88	9.87	-0.13	0.01	0.47	
		9.88	9.88	9.87	9.86	9.87					
2	50.0	49.80	49.80	49.82	49.81	49.85	49.81	-0.19	0.02	0.47	
		49.82	49.80	49.82	49.82	49.78					
3	100.0	99.84	99.86	99.87	99.87	99.88	99.87	-0.13	0.01	0.47	
		99.87	99.88	99.86	99.87	99.88					

**Remarks**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,  (Calibration Engineer) <b>M.RAGUL</b>	Authorised by:  (Quality Manager/Chief Executive) <b>C.SIVABALAN</b>
--	---





## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13881

Page No: 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
MECHERI-636453,SALEM DISTRICT.

SRF No.	SRF/23/00831-0004
SRF Date	03-10-2023
Date of Receipt	03-10-2023
Date of Calibration	03-10-2023
Due Date for Calibration	02-10-2024
Issue Date	03-10-2023

**Details of Unit Under Calibration**

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW11507	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

**Environmental Condition**

**Calibration Method Used**

Temperature (°C)	23.2	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

**Result of Calibration in µl**


SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.94	99.95	99.92	99.96	99.94	99.94	-0.06	0.02	0.47
		99.91	99.92	99.94	99.95	99.96				
2	500	499.85	499.87	499.88	499.87	499.89	499.88	-0.12	0.01	0.47
		499.87	499.89	499.90	499.88	499.89				
3	1000	999.74	999.76	999.78	999.77	999.75	999.77	-0.23	0.02	0.47
		999.75	999.77	999.78	999.77	999.79				

**Remarks**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)  
M.RAGUL



Authorised by:

(Quality Manager/Chief Executive)  
C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.