

		CALIE	BRATIO	N CER	TIFICA	TE		
CERTIFICATE NO: SBS/CL/23/13596			MECHANICA		AL	Page No:	1 of 1	
Issue Date			07-10-2023					
SRF No & Date			SRF/23/00775-0002 & 06-10-2023					
Receipt Date			06-10-203					
Calibration Date			06-10-2023					
Calibration Due			05-10-2024					
Custo	mer Name & Add	ress						
GOVE	RNMENT PRIMA	ARY HEALTH CENTRE						
AYAK	ARANPULAM-61	4707, NAGAPATTINAM DIS	STRICT.					
		Deta	ils of Device	Under Calib	ration (DUC)	)		
Descr	iption	CENTRIFUGE	Mak	e & Model	:	LABOTECH	& BDI-152	
Range : 3500 RPM		Sr. No		:	NA			
Resolution 1 RPM		Identification No :			NA			
DUC Condition : Satisfactory			Location			LABORATO	RY	
		Environmental Co	onditions & S	tandard Ope	erating Proc	edure Details		=L==-KEe+-
Environmental Details Temperature			e: 25.5°C Relative Hu		midity 52%Rh			
Calibration Procedure No SBS/CP/ML/			'04 Calibration of		done at ONSITE			
			Reference	Standards D	etails			
S.No	Description		Make/ SI No:		Certificate No			Validity
1	1 Digital Tachometer		LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		13	09-06-2024

## **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	500	499.0	1.0	4.2	
2	1500	1499.5	0.5	4.2	
3	3500	3499.5	0.5	4.2	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards,

Calibration Engineer)
P.PRASANNA



	77.77 R. M.
(Quality Manager)	(Chief Executive
(C.SIVABALAN)	

Authorised Signatory