

			CAL	BRAT	101	CER	TIFIC	A'	TE_	
CERTIFICATE NO: SBS/CL/23/13597					MEDICAL DEVICES					Page No:1 of 1
Issue Date					07-10-2023					
SRF No & Date				SRF/23/00775-0003 & 06-10-2023						
Receipt Date					06-10-203					
Calibration Date					06-10-2023					
Calibration Due					05-10-2024					
Cust	omer Name & Addre	55								
GOV	ERNMENT PRIMA	RY HEA	LTH CENTRE							
AYA	KARANPULAM-614	707,NA	GAPATTINAN	DISTRICT						
				Details of D	evice U	nder Calibra	tion (DUC	;)		
		ELEC1	RICAL							
Description		SAFETY(MICROSCOPE)			Make & Model		1		LABOMED & CXL PLUS	3
Range		MULTI			Sr. No				130862608	
Resolution		MULTI		Identification No		:		NA		
DUC Condition		SATISFACTORY		Location				LABORATORY		
			Environn	nental Cond	litions	& Calibration	Procedu	re C	Details	
Environmental Details		Temperature:25.6°C		25.6°C	Relative Humi		midity	nidity 52% RH		
Calibration Procedure No		SBS/CP/MD/29		Calibration do		done at ONSITE		ONSITE		
				Refer	ence S	tandards De	tails			
S.No	Description			Make/ SI No:		Certificate		ate l	No	Validity
1	Electrical Safety Analyser R		Rigel Medical & 44L-1059		M-230809-16-4			10-08-202		

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)	
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)	
	>20MΩ	83	13.92	
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)	
	<5000µAfor B,BF,CF	181	9.7	
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)	
	<500µAfor B,BF,CF	231	8.5	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

(Calibration Engineer)
P.PRASANNA



-	Technical Manager
	(C.SHANMUGARAJ)

Chief Executive