

## CHRISTIAN MEDICAL COLLEGE

## **DEPARTMENT OF CLINICAL BIOCHEMISTRY**

## **CMC EXTERNAL QUALITY ASSURANCE SCHEME**





**Lab Name** Govt. PHC, Ananthur Lab No 12332

Constituent Date of Result Entered: 19/06/2023 **Chemistry III** Group

PT item Lyophilized human serum based Date of Report Published: 04/07/2023

CLNa	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	
SI.No						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	306 mg/dL	1.31	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	112 mg/dL	-0.34	0.24

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Ananthur THE MEDICAL OFFICER, Mathur, Krishnagiri **Ananthur** TAMILNADU635304

> **Coordinator Contact Details:** Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss **CMC EQAS Coordinator Christian Medical College, Vellore** 

Panela Christudoss

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components \*\*\*\*\*\* End of Report \*\*\*\*\*