

	CALIBRATI				
CERTIFICATE NO: SBS/CI	/23/10354	ME	CHANICAL	Page	e No:1 of 1
Issue Date	18	18-08-2023			
SRF No & Date	SF	SRF/23/00319-0003 & 17-08-2023			
Receipt Date	17	17-08-2023			
Calibration Date	17	17-08-2023			
Calibration Due	16	16-08-2024			
Customer Name & Address				What has a little and the	
GOVERNMENT PRIMARY HEALT	H CENTRE,				
ANANDUR-635304, KRISHNAGIR					
	Details of Devi	ce Under Calibra	tion (DUC)		
Description : CENTR	IFUGE Ma	Make & Model		LABOTECH & NA	
Range : 34	00 RPM Sr	. No	:	9328	
Resolution :	1 RPM	entification No	:	NA	
DUC Condition : Satisfac	tory Lo	Location		LABORATORY	
E	nvironmental Conditions	& Standard Oper	ating Procedu	ire Details	
Environmental Details	Temperature: 24.3 °C				
Calibration Procedure No	SBS/CP/ML/04			done at ONSITE	
- Carlotte Control of the Control of	Referen	ce Standards De	tails		
S.No Description	Make/ SI N	Make/ SI No:		Certificate No	
	Lutron & C	Lutron & Q638380		JRPM-CCTR-A&S-2022-0036	
Digital Tachometer	Lutron & C	Lulion & Q030300		.,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
4	1000	999.5	0.5	4.2
2	2500	2499.7	0.3	4.2
3	3400	3399.5	0.5	4.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) C.SIVABALAN



(D.VETRI SELVI)

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

