

LAB MONTHLY SUMMARY

**Lab Name** Govt. PHC, A.Vellalapati [Details About Robust Analysis](#)  
**Lab No** 12367 [Detail About Monthly Summary](#)  
**Month** July [Detail about SDI](#)  
**Year** 2023  
**Constituent** Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/07/2023


Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	24	109.83	6.83	7.50	99 mg/dL	-1.44	3.06
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	18	5.79	3.00	0.17	5.8 mg/dL	0.06	0.08
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	112.64	4.38	4.93	118 mg/dL	1.09	2.26

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
 L.T

## W LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, A.Vellalapatti	Details About Robust Analysis
Lab No	12367	Detail About Monthly Summary
Month	June	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	22	283.77	5.80	16.44	294 mg/dL	0.62	7.01
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	23	2.50	9.65	0.24	2.2 mg/dL	-1.24	0.11
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	22	114.61	5.43	6.22	111 mg/dL	-0.58	2.61

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

[http://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2\\_new.aspx?resDetId=266&noSamples=1&month=](http://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=266&noSamples=1&month=)

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, A.Vellalapatti	Details About Robust A
Lab No	12367	Detail About Monthly S
Month	May	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

Date of Result Entered : 20

Date of Report Published : 0

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	177.53	5.62	9.98	195 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	22	1.85	7.73	0.14	1.9 mg/dL
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	20	115.73	3.06	3.54	118 mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>





**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name Govt. PHC, A.Vellalapatti Lab No 12367  
Constituent Group Chemistry III Date of Result Entered : 21/04/23  
PT item Lyophilized human serum based Date of Report Published : 06/05/23

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	118.42	8.48	10.04	108 mg/dL	-1.04
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	19	2.14	6.03	0.13	2.3 mg/dL	1.24
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	20	111.15	4.25	4.72	114 mg/dL	0.60

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, A.Vellalapatti  
CLINICAL LABORATORY  
A.Vellalapatti  
TAMILNADU625301

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
21/5

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, A.Vellalapatti	Details About Robust Analysis
Lab No	12367	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	23	173.17	4.83	8.36	170 mg/dL	-0.38	3.48
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.74	6.34	0.11	1.7 mg/dL	-0.36	0.01
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	17	111.42	3.18	3.54	113 mg/dL	0.45	1.7

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*dm*  
1.1

VIEW LAB MONTHLY SUMMARY

**Lab Name** Govt. PHC, A.Vellalapatti Details About Robust Analysis  
**Lab No** 12367 Detail About Monthly Summary  
**Month** September Detail about SDI  
**Year** 2023  
**Constituent Group** Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 21/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	18	307.20	6.79	20.85	298 mg/dL	-0.44	9.83
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.26	9.52	0.12	1.2 mg/dL	-0.50	0.05
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	113.64	2.93	3.33	113 mg/dL	-0.19	1.53

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*Handwritten signature*  
L.S