



## CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13960	MECHANICAL	Page No:1 of 1	
Issue Date	10-10-2023			
SRF No & Date	SRF/23/00844-0002 & 09-10-2023			
Receipt Date	09-10-2023			
Calibration Date	09-10-2023			
Calibration Due	08-10-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, KOORAM-631552,KANCHIPURAM DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	CENTRIFUGE	Make & Model	REMI & R-8C	
Range	5250 RPM	Sr. No	ZCIN-27315	
Resolution	10 RPM	Identification No	NA	
DUC Condition	Satisfactory	Location	LABORATORY	
<b>Environmental Conditions &amp; Standard Operating Procedure Details</b>				
Environmental Details	Temperature: 25.4°C	Relative Humidity	54% Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

### **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	998.6	1.4	4.2
2	2000	1998.6	1.4	4.2
3	3000	2998.7	1.3	4.2

### **REMARKS**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor  $k=2.00$ .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

*KSPM*  
(Calibration Engineer)  
K.SATHYAMOORTHY



Authorised Signatory

*C. Sivabalan*  
Quality Manager  
(C.SIVABALAN)       Chief Executive

## **SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.